





---

## Microscope Objectives ∞ Phase Contrast

$\infty$ Phase Contrast				
Magnification	4X	10X	40X	100X (Oil)
Numerical Aperture	0.1	0.25	0.65	1.25
Working Distance (mm)	16.8	5.8	0.43	0.13
Slide CT (mm)	-	0.17	0.17	0.17
BFL (mm)	61.4004	-10.6771	-21.4309	-22.1035
Conjugate Distance (mm)	$\infty$	$\infty$	$\infty$	$\infty$
Wavelength (nm)	3.355	1.342	0.516	0.268
Resolution ( $\mu\text{m}$ )	63.214	10.114	1.2	0.29

 : [www.shanghai-optics.com](http://www.shanghai-optics.com)

 : [rfq@shanghai-optics.com](mailto:rfq@shanghai-optics.com)

 : 732-321-6915